

Interactive comment on “Background subtraction for the Cluster/CODIF plasma ion mass spectrometer” by C. G. Mouikis et al.

C. G. Mouikis et al.

chris.mouikis@unh.edu

Received and published: 6 January 2014

A table (Table 1) that shows the cross species contamination from pre-flight calibration data is included. In addition, in Section 2 the following paragraph is included:

"In fact, the tails of the distribution of each species "spill" into the other species at a certain level. Table 1 shows the spillover fraction of the different species, for the two energies shown in Figure 2, from pre-flight calibration data obtained from the high sensitivity side of S/C4. In order to assess the significance of the contamination level between species, the spillover fraction and the relative species abundance has to be considered. For example, inside the magnetosphere the He⁺⁺ is almost irretrievable because the spillover from H⁺ (~1.1-5.1%) is large compared to the real He⁺⁺ signal.

C233

In contrast, the H⁺ spillover into the O⁺ TOF window is much lower (~0.1-0.3%) while the real O⁺ counts are much higher inside the magnetosphere, which results in a much lower contamination level."

Interactive comment on Geosci. Instrum. Method. Data Syst. Discuss., 3, 567, 2013.