

Interactive comment on “FTS measurements of column CO₂ at Sodankylä” by R. Kivi and P. Heikkinen

Anonymous Referee #3

Received and published: 25 February 2016

The manuscript "FTS measurements of column CO₂ at Sodankylä" by Kivi and Heikkinen describes the setup of the TCCON Fourier Transform Spectrometer at the Finnish Meteorological Institute in Sodankylä, Finland. Setup and history of the instrument and site are well documented.

The paper will certainly be useful as a site reference paper for future publications. It does not contain a deep analysis of the data and only describes the XCO₂ time series and some - certainly interesting - quality checking parameters. In fact, it would be good to have similar publications for other TCCON sites as well.

In its current form, the article should be appropriate for the scope of Geosci. Instrum. Method. Data Syst. It would, however, be more interesting if

a) the XCO₂ time series was compared to some other data series - even not in very

[Printer-friendly version](#)

[Discussion paper](#)



deep detail

b) other time series of measured trace gases like XCH₄, XCO, XN₂O, XH₂O could at least be shown in an overview - as I do not think the authors plan to submit similar articles for these species.

General comments:

- please make sure to use CO₂ vs. XCO₂ vs. column CO₂ correctly throughout the text
- the use of tense is inconsistent: very often present tense is used for descriptions of past events.

Specific comments:

Title and abstract: FTS actually measures XCO₂ instead of CO₂

p. 1, l. 26: XCO₂: "column averaged dry-air mole fraction"

p. 2, l. 14: start a new sentence after "in early 2009".

p. 3, l. 6: I guess 45 cm OPD is the typical value and not the maximum - as "up to" implies.

p. 3, l. 7: rather talk about XCO₂, XCH₄, ...

p. 4, l. 6: given that you have easy access to your instrument: why are there so few ILS measurements?

Fig. 1: better also label x-axis of top plot (even though it is the same as the lower)

Fig. 2: please mark the time when the laser was changed

Fig. 4: I am confused by the y axis: how can you have up to 600 spectra per day? At 10 kHz scanner velocity, a full forward-backward scan should take around 3-4 minutes.

Interactive comment on Geosci. Instrum. Method. Data Syst. Discuss., doi:10.5194/gi-2015-

[Printer-friendly version](#)

[Discussion paper](#)



[Interactive
comment](#)

[Printer-friendly version](#)

[Discussion paper](#)

