

Interactive comment on “Development of high-precision distributed wireless microseismic acquisition stations” by Shuaiqing Qiao et al.

Anonymous Referee #2

Received and published: 13 July 2018

1. The seismic test system has used ARM the master controller, why use the FPGA module ?
2. Line68, mentioned that the differential signal from the detector is converted into a single-ended signal through the amplifier, and then the single-ended signal is converted into a differential signal and converted into a differential signal input to the analog-to-digital conversion chip, why?
3. Chapter 4 mentions the connection between the central station and each relay, issuing commands and real-time data recovery. The data collected here means that all data collected by the collection station is transmitted by wireless.
4. In the conclusion of Section 5.1, the "equivalent error" is $0.51\mu V$, which should be

[Printer-friendly version](#)

[Discussion paper](#)



"equivalent input noise" and the expression is incorrect.

Interactive comment on Geosci. Instrum. Method. Data Syst. Discuss.,
<https://doi.org/10.5194/gi-2018-17>, 2018.

GID

[Interactive
comment](#)

[Printer-friendly version](#)

[Discussion paper](#)

