

Reply to referee comments

Manuscript ID: GI-2020-31

Error estimate for fluxgate magnetometer in-flight calibration on a spinning spacecraft

Y. Narita, F. Plaschke, W. Magnes, D. Fischer, and D. Schmid

Referee 1

1. 1. General remarks

The submitted manuscript provides an analysis of the errors of the FGM magnetic field data related to the uncertainties of the calibration parameters determined for spinning spacecraft through the procedure described by Plaschke et al. 2019. The errors are computed in the first order by introducing small perturbations of the calibration parameters and taking into account typical values of the uncertainties as determined for a number of spacecraft (Cluster, THEMIS, MMS, and BepiColombo).

The manuscript can be a useful resource both for planning future spacecraft missions and for estimating the accuracy of the magnetic field data delivered by current missions. There are however a number of issues which should be addressed before publication.

Reply:

- Thank you for careful reading, thoughtful comments, and positive evaluation.

2. 2. Specific comments

page 2, line 29–30: Please explain what do you mean through “Fourier series of spacecraft spin frequency”.

Reply:

- We changed the text as follows to avoid confusion.

- page 2, line 31–35
“...Fourier series over the frequencies as

$$B_i(t) = \sum_{n=0}^{N-1} F_i(\omega) e^{in\omega t}$$

for the i -th component of magnetic field. F_i is the Fourier coefficient, i the imaginary unit, ω the de-spinning frequency (as angular frequency), N the number of data points, and t the time in the data.”

3. page 2, line 45–49: *Is ϕ_a the angle between the projection of the spin axis on the (Px, Py) -plane of the sensor package coordinate system? If so, please mention this in the description of the step 3.*

Reply:

- No, ϕ_a is the azimuthal angle around the spacecraft spin axis in the xy-plane in the coord-2 system.
- page 3, line 56–59
“first by rotating ... (with the rotation angle ϕ_a in the xy-plane around the spin axis in the coord-2 system) and then by orienting...”

4. page 2, line 48: *“... σ_{Px} and σ_{Py} (with respect to the Pz axis) ...” This definition of σ_{Px} and σ_{Py} is not clear enough. From Fig. 1 it appears that σ_{Py} is the angle between the Pz axis and the projection of the spin axis on the (Pz, Py) -plane. σ_{Px} appears to be the angle between the spin axis and the (Pz, Py) -plane. If this is correct, please clarify this in the description of step 3.*

Reply:

- The definition of the angles σ_{Px} and σ_{Py} suggested by the referee is perfect and unambiguous. Thank you for the suggestion. We included the definition.
- page 3, line 61–62
“Here, σ_{Py} is the angle between the Pz axis and the projection of the spin axis on the (Pz, Py) -plane. σ_{Px} is the angle between the spin axis and the (Pz, Py) -plane.”

5. page 5, line 55–59: *The description of step 5 is misleading. A clear distinction should be made between the ideal exact transformations and the transformations derived through the calibration procedure. The “above transformations” as described in steps 1–4 are the exact transformations (not derived through calibration, thus not affected by calibration errors). If they are inverted then the true field is obtained. In lines 56–59 the Authors refer to the transformations estimated through the calibration procedure. To eliminate the potential confusion I suggest that at each step the corresponding transformation is explicitly stated. E.g. step 2: coord 1 Ω^{-1} coord 2; step 3: coord 2 $\Sigma^{-1}\Phi^{-1}$ coord 3; step 4: coord 3 $G^{-1}\Gamma^{-1} + O_s$ coord 4; step 5: coord 4 $\Omega'\Phi'\Sigma'\Gamma'G' - O_s^{\text{prime}}$ coord 5; direct transform: coord 4 $\Omega\Phi\Sigma\Gamma G - O_s$ coord 1; The prime symbol denotes the quantities derived using the calibration procedure. Alternatively, the transformations could be added to Fig. 1.*

Reply:

- Good idea. We added equations in each step to make the explanations unambiguous.

- page 2, line 43–44

“... The magnetic field modeled in the coord-1 system as

$$\mathbf{B}_{c1} = \begin{bmatrix} B_X \\ B_Y \\ B_Z \end{bmatrix} = \begin{bmatrix} B_p \\ 0 \\ B_a \end{bmatrix}$$

...”

- page 3, line 53–55

“...The magnetic field vector in the coord-2 system is symbolically related to that in the coord-1 system as

$$\mathbf{B}_{c2} = \mathbf{\Omega}^{-1} \mathbf{B}_{c1},$$

where $\mathbf{\Omega}^{-1}$ is the spin rotation matrix. Note that $\mathbf{\Omega}$ is defined as the de-spinning matrix here.”

- page 3, line 62–67

“...The magnetic field vector in the coord-3 system is symbolically related to that in the coord-2 system as

$$\mathbf{B}_{c3} = \mathbf{\Sigma}^{-1} \mathbf{\Phi}^{-1} \mathbf{B}_{c2},$$

where $\mathbf{\Phi}^{-1}$ is the azimuthal rotation matrix in the spin plane (around the spin axis in the coord-2 system) and $\mathbf{\Sigma}^{-1}$ is the transformation matrix to orient the z axis in the direction to the sensor package Pz direction. Again the matrices without inversion are used for the reconstruction of the model magnetic field in the calibration.”

- page 3, line 72–76

“...The magnetic field vector in the coord-4 system is symbolically related to that in the coord-3 system as

$$\mathbf{B}_{c4} = \mathbf{G}^{-1} \mathbf{\Gamma}^{-1} \mathbf{B}_{c3} + \mathbf{O}_s,$$

where $\mathbf{\Gamma}^{-1}$ is the transformation matrix using three angles (θ_1 , θ_2 , and ϕ_{12}), \mathbf{G}^{-1} is the gain matrix, and \mathbf{O}_s is the offset vector.”

6. page 3, line 55, 60–61: *While correct, the formulations in lines 55 “... transformations are inverted to estimate the magnetic field ...” taken together with the lines 60-61, “... the forward transform is defined for the conversion ...” can be confusing for the reader. Adding the transforms to the steps as suggested above and reformulating line 55 could help clarifying this.*

Reply:

- We added an equation for the reconstruction of the model field using the sensor magnetic field and the calibration matrices.

- page 4, line 81–84
“... The model magnetic field is reconstructed from the sensor magnetic field as

$$\mathbf{B}_{c5} = \mathbf{\Omega} \mathbf{\Phi} \mathbf{\Sigma} \mathbf{\Gamma} \mathbf{G} (\mathbf{B}_{c4} - \mathbf{O}_s).$$

If the calibration parameters are all known, the reconstructed field \mathbf{B}_{c5} restores the model field \mathbf{B}_{c1} .”

- page 4, line 89–92
“...The relation between the sensor-output magnetic field $\mathbf{B}_s = \mathbf{B}_{c4}$ (introduced in the coord-4 system) and the model ambient field in the spinning frame \mathbf{B}_{c2} (introduced in the coord-2 system, Eqs. 3–5) is expressed by a set of transformation matrices $\mathbf{G}^{-1} \mathbf{\Gamma}^{-1} \mathbf{\Sigma}^{-1} \mathbf{\Phi}^{-1}$ and an offset vector \mathbf{O}_s as (Plaschke et al., 2019)

$$\mathbf{B}_s = \mathbf{G}^{-1} \mathbf{\Gamma}^{-1} \mathbf{\Sigma}^{-1} \mathbf{\Phi}^{-1} \mathbf{B}_{c2} + \mathbf{O}_s.$$

...”

7. page 4, line 82: *“(angle between the coord 2 system and the coord 3 system)” formulation is ambiguous. Please reformulate.*

Reply:

- We changed the text as follows.

- page 5, line 108–110
“...in sensor package system (σ_{Py} is the angle between the sensor-3 direction and the projection of the spin axis onto the sensor package Py–Pz plane; σ_{Px} is the angle of spin axis and the sensor package Py–Pz plane)”

8. page 4, line 84–86: *The descriptions of $\delta\theta_1$, $\delta\theta_2$ and $\delta\Phi_{12}$ are unclear. Perhaps instead of “with a relation to the deviation from 90° ”, simply writing “the deviation from 90° of the elevation/azimuth” angles would be better.*

Reply:

- Rewritten as follows.

 - page 5, line 111
“deviation of elevation angles from 90° defined as $\delta\theta_1$ and $\delta\theta_2$, for the sensors 1 and 2, respectively”

 - page 5, line 112
“deviation of azimuthal angle from 90° defined as $\delta\phi_{12}$ ”
9. *page 5, line 115–118: The formulation is difficult to understand. Please reformulate to make clear that the coord 5 system is obtained using the transformation matrices resulted from the calibration procedure.*

Reply:

- We reformulated as follows.

 - page 6, line 141–143
“Here, the magnetic field vector ($B_{X'}$, $B_{Y'}$, $B_{Z'}$) is represented in the coord-5 system and hence ideally reproduce the model magnetic field in the coord-1 system. That is, the z-component is in the direction of spacecraft spin axis, the x-component is in the spin plane. The y-component is also in the spin plane but should ideally not contain the ambient field.”
10. *page 12, section 4: It might be useful to rewrite the estimated errors (Eqs. 27, 30, 33) using the expected values given in Table 1.*

$$\begin{aligned}
 |\Delta B_{x'}| &\leq 0.1nT + (B_p + B_a) \times 10^{-2} \\
 |\Delta B_{y'}| &\leq 0.1nT + (10B_p + B_a) \times 10^{-3} \\
 |\Delta B_{z'}| &\leq 0.21nT + (B_p + B_a) \times 10^{-3}
 \end{aligned}$$

or similar.

Reply:

- Good idea!

- page 13, line 300–305
“... For a practical purpose, the combined errors in Eqs. (33), (36), and (39) are reformulated in an approximate form using the values given in

Tab. 1:

$$\begin{aligned} |\Delta B_{x'}| &\leq 0.1 \text{ [nT]} + (B_p + B_a) \times 10^{-2} \\ |\Delta B_{y'}| &\leq 0.1 \text{ [nT]} + (10B_p + B_a) \times 10^{-3} \\ |\Delta B_{z'}| &\leq 0.2 \text{ [nT]} + (B_p + B_a) \times 10^{-3}. \end{aligned}$$

...”

11. *page 13, 14, Fig 2, 3: Many spacecraft measure magnetic fields much higher than 300 nT. Please extend the ambient magnetic field domain (x-axis) to 10 000 nT.*

Reply:

- The plots were extended to ambient field to 10,000 nT and were added as Fig. 4 and Fig. 5. We still keep the original low-field case (Fig. 2 and Fig. 3, up to 300 nT) for the potential use to the BepiColombo Mio spacecraft.
- Figures 4 and 5 (pages 15 and 16).

12. *page 14, lines 303–304: “... when the ambient field is aligned with the spin axis ...”. From Eq.(30) Φ_a multiplies B_p , therefore the largest error due to Φ_a is for the ambient magnetic field orthogonal to the spin axis.*

Reply:

- True! Thank you!
- page 16, line 336
“...ambient field is aligned in the spin plane.”

13. *3. Minor points*

page 1, line 5 and elsewhere in the manuscript: “... perturbing the calibration procedure ..”. — not the procedure, but the calibration parameters are perturbed.

Reply:

- Right! Changed into “calibration parameters.” (page 1, line 5; page 16, line 337).

14. *page 1, line 7–8: “The final error ... are important factors.” this sentence is difficult to understand. Please reformulate.*

Reply:

- The sentence was simplified.
- page 1, line 7–8
“The final error depends on (1) the magnitude of magnetic field with respect to the offset error and (2) the angle of magnetic field to the spacecraft spin axis.”

15. *page 2, line 31: “sensor output”, “not ourput”*

Reply:

- Done (page 2, line 36).

16. *page 2, line 32: “sensor-axes” (if plural)*

Reply:

- Done. (page 2, line 37; page 5, line 98)

17. *page 4, line 72: “The”, not “There”*

Reply:

- Done. “The matrices” on page 5, line 98. It was a misspelling of “These” but I find “The” also appropriate.

18. *page 4, line 82: remove “to the angles”*

Reply:

- Done (page 5, line 108).

19. *page 13, 14, Fig 2, 3: The gray lines cannot be distinguished from the black lines. Please use colors or solid/dashed lines.*

Reply:

- The gray curve is re-plotted with dashed line in darker, thicker gray in the revision (Fig. 2 and Fig. 3, pages 14 and 15).

20. *page 14, line 307: remove one “how”*

Reply:

- Done. “shows how” (page 16, line 339).
-

Referee 2

1. *Full disclosure: this reviewer also reviewed Plaschke et al 2019.*

While starting this review I had some concern there would be a large overlap with the earlier paper, and that my eyes would not be “fresh” enough. Happily this is not the case. The review of the earlier paper has the right level of detail, enough to provide a good picture without being excessive. The first reviewer provides a good discussion of this part of the present manuscript so I leave it at that.

I agree with the first reviewer, that the grey line/black line markings in Figs 2 and 3 are difficult to observe.

Finally, it appears that this study of errors is targeted at low-field measurement settings. If this is the case it should be made clear. There is also no discussion of magnetometer nonlinearity errors as distinct from the errors due to the nonlinear coupling of the parameters under consideration. The latter is suitably noted in the paper, and in low-field settings sensor non linearity is usually negligible.

In low earth orbit this can be different. Modern sensors which are often double wound, and even triple wound have excellent linearity, but not always. The 1979 MAGSAT single-wound sensor suffered from about 1% nonlinearity, and the same sensor design was used more recently on MESSENGER. With present thinking about the possibility of deploying large fleets of small magnetometer cubesats with just as small sensors one might ask whether nonlinearity issues can rise again. Also the potential quantity of spacecraft points to wanting automated calibration and error estimation methods. Perhaps the methods of Plaschke et al 2019 and the present paper will make that task easier.

Reply:

- Figs. 2 and 3 were improved using darker, thicker, dashed lines (page 14, page 15).
- We address the scope of manuscript more clearly in the introduction.
- page 1, line 23 to page 2, line 26
“The scope of our work is the error estimate of calibrated magnetometer data in a low-field environment. In practice, more effects need to be taken into account, including sensor nonlinearities, temperature dependence (temperature drift effect), and jumps in the data associated with the change in operational modes.”

- We agree with the importance of nonlinearity issues. We added the following sentences at the end of manuscript.

- page 16, line 349 to page 17, line 356

“The errors associated with the uncertainties in calibration parameters are studied in this paper. In a low-field environment such as in interplanetary space the sensor nonlinearity (which originates in the nonlinearity of gain) is usually considered negligible. In a low Earth orbit the situation may be different. Modern sensors which are often double wound, and even triple wound have excellent linearity (typically to an accuracy of about 10^{-4} per axis), but this is not always the case. The MAGSAT single-wound sensor (Acuña, 1980; Langel et al., 1982), for example, suffered from about 1% nonlinearity, and the same sensor design was used more recently on MESSENGER (Solomon et al., 2007; Anderson et al., 2007). With present thinking about the possibility of deploying large fleets of small magnetometer cubesats with just as small sensors one might ask whether nonlinearity issues can rise again.”

2. *Here is a small collection of additional minor corrections:
2–31 “sensor ourput”*

Reply:

- Corrected to “sensor output” (page 2, line 36).

3. *9–189 “seconr order”*

Reply:

- Corrected to “second order” (page 9, line 212).

4. *9–206 “?Auster”*

Reply:

- Corrected to “THEMIS (Angelopoulos, 2008; Auster et al., 2008)” (page 10, line 231).

5. *12–266 “accuracy fo”*

Reply:

- Corrected to “accuracy of” (page 13, line 291).

6. 12-277 “Figure 2)”

Reply:

- Corrected to “Fig. 2” (page 13, line 307).

7. 14-307 “sows how how”

Reply:

- Corrected to “show how” (page 16, line 339).

8. 14-311 “error throughly”

Reply:

- Corrected to “error thoroughly” (page 16, line 343).

Other changes

The following literatures were added:

- Acuña, M. H.: The MAGSAT precision vector magnetometer, Johns Hopkins APL Technical Digest, 1, 3, 210213, 1980.
- Anderson, B. J., Acuña, M. H., Lohr, D. A., Scheifele, J., Raval, A., Korth, H., and Slavin, J. A.: The magnetometer instrument on MESSENGER, Space Sci. Rev., 131, 417450, 2007. <https://doi.org/10.1007/s11214-007-9246-7>
- Langel R., Ousley G., Berbert J., Murphy J., and Settle M.: The MAGSAT Mission. Geophys. Res. Lett., 9, 243245, 1982. <https://doi.org/10.1029/GL009i004p00243>
- Solomon, S. C., McNutt Jr., R. L., Gold, R. E., Domingue, D. L.: MESSENGER mission overview, Space Sci. Rev., 131, 339, 2007. <https://doi.org/10.1007/s11214-007-9247-6>

Error estimate for fluxgate magnetometer in-flight calibration on a spinning spacecraft

Yasuhito Narita¹, Ferdinand Plaschke¹, Werner Magnes¹, David Fischer¹, and Daniel Schmid¹

¹Space Research Institute, Austrian Academy of Sciences, Schmiedlstr. 6, A-8042 Graz, Austria

Correspondence: Y. Narita
(yasuhito.narita@oeaw.ac.at)

Abstract. Fluxgate magnetometers are widely used for in-situ magnetic field measurements in the context of geophysical and solar system studies. Like in most of experimental studies, magnetic field measurements using the fluxgate magnetometers are constrained to the associated uncertainties. To evaluate the performance of magnetometers, the measurement uncertainties of calibrated magnetic field data are quantitatively studied for a spinning spacecraft. The uncertainties are derived analytically by [perturbing the calibration parameters](#), and are simplified into the first-order expression including the offset errors and the coupling of calibration parameter errors with the ambient magnetic field. The error study shows how the uncertainty sources combine through the calibration process. [The final error depends on \(1\) the magnitude of magnetic field with respect to the offset error and \(2\) the angle of magnetic field to the spacecraft spin axis.](#) The offset uncertainties are the major factor in a low-field environment, while the angle uncertainties (rotation angle in the spin plane, sensor non-orthogonality, and sensor misalignment to the spacecraft reference directions) become more important in a high-field environment in a proportional way to the magnetic field. The error formulas serve as a useful tool in designing high-precision magnetometers in future spacecraft missions as well as in data analysis methods in geophysical and solar system science.

1 Introduction

Fluxgate magnetometers perform measurements from DC (direct current) to low-frequency magnetic field vectors (typically up to 10–100 Hz), and are widely applied to in situ spacecraft observations for space plasma, magnetospheric, and heliospheric research (Acuña, 2002). The fluxgate magnetometers can be mounted on a spinning spacecraft or three-axis stabilized one, depending on the individual mission concept. In particular, in-flight calibration benefits from the spacecraft spin, since 8 of 12 calibration parameters are determined by making use of the spacecraft spin. Detailed procedure for the in-flight calibration on a spinning spacecraft are presented by, e.g., Kepko et al. (1996) and Plaschke et al. (2019).

The goal of the current paper is to give an outline of systematic errors of calibrated fluxgate magnetometer data on a spinning spacecraft. The error of magnetic field data occurs due to the uncertainties of the calibration parameters. The error sources may combine with one another through the calibration process. We derive the full expression of calibration errors as well as a more practical, simplified expression by truncating at the first order of relative errors. [The scope of our work is the error estimate of calibrated magnetometer data in a low-field environment. In practice, more effects need to be taken into account, including](#)

25 sensor nonlinearities, temperature dependence (temperature drift effect), and jumps in the data associated with the change in operational modes.

2 Systematic error on in-flight calibration

For a spin-stabilized spacecraft, the magnetometer in-flight calibration is performed by correcting for offsets (including the spacecraft DC field), gains, deviations from the ideal orthogonal coordinate system, spacecraft spin axis direction with respect
 30 to the sensor reference direction and rotation angle around the spacecraft spin axis. For a nearly-orthogonal unit-gain sensor system, the measured magnetic field is transformed into a de-spun coordinate system, and is expanded into a [Fourier series](#) over the frequencies as

$$B_i(t) = \sum_{n=0}^{N-1} F_i(\omega) e^{in\omega t} \quad (1)$$

for the i -th component of magnetic field. F_i is the Fourier coefficient, i the imaginary unit, ω the de-spinning frequency (as
 35 angular frequency), N the number of data points, and t the time in the data.

The magnetic field vector measured by the three sensors ([sensor output](#)) is related to the ambient field by taking account of spacecraft spin-axis direction, spacecraft spin phase, [sensor-axes directions](#), sensitivities (or gains) of the sensors, and offsets (Kepko et al., 1996; Plaschke et al., 2019). The relation is constructed in the following fashion.

1. The true or model ambient field is set in the inertial (i.e., non-spinning) orthogonal spacecraft spin axis-aligned coordinate system (the coord-1 system in Fig. 1) with the spin-plane component in the X direction ($B_X = B_p$) and the
 40 spin-axis component in the Z direction ($B_Z = B_a$). There is no magnetic field in the rest spin-plane component, $B_Y = 0$, because the coord-1 system spans the spacecraft spin axis (in the Z direction) and the ambient field in the X-Z plane.

The magnetic field modeled in the coord-1 system as

$$\mathbf{B}_{c1} = \begin{bmatrix} B_X \\ B_Y \\ B_Z \end{bmatrix} = \begin{bmatrix} B_p \\ 0 \\ B_a \end{bmatrix}. \quad (2)$$

45

2. The model ambient field in the coord-1 system is transformed into the spinning orthogonal spin-axis-aligned system (the coord-2 system in Fig. 1) with the magnetic field components B_x , B_y , and B_z by referring to the spin axis as the z direction and rotating the spin plane around the spin axis by the spacecraft spin phase $-\omega t$ (here ω is defined as the de-spinning frequency and $-\omega$ as the spin frequency; t the time) as

$$50 \quad B_x = B_X \cos(-\omega t) \quad (3)$$

$$B_y = B_Y \sin(-\omega t) \quad (4)$$

$$B_z = B_Z. \quad (5)$$

The magnetic field vector in the coord-2 system is symbolically related to that in the coord-1 system as

$$B_{c2} = \Omega^{-1} B_{c1}, \quad (6)$$

55 where Ω^{-1} is the spin rotation matrix Note that Ω is defined as the de-spinning matrix here.

3. The field is then transformed into the spinning, orthogonal sensor package system (the coord-3 system in Fig. 1) first by rotating around the spin axis by correcting for the magnetometer boom extension and a possible misalignment of the fluxgate sensor in the spin plane (with the rotation angle ϕ_a in the xy -plane around the spin axis in the coord-2 system) and then by orienting the Pz axis in the sensor-3 direction with the spin axis tilt angles σ_{Px} and σ_{Py} (with respect to the Pz axis) to obtain the magnetic field components as B_{Px} , B_{Py} , and B_{Pz} (here, P in the subscript stands for the sensor package). Here, σ_{Py} is the angle between the Pz axis and the projection of the spin axis on the (Pz, Py) -plane. σ_{Px} is the angle between the spin axis and the (Pz, Py) -plane. The magnetic field vector in the coord-3 system is symbolically related to that in the coord-2 system as

$$B_{c3} = \Sigma^{-1} \Phi^{-1} B_{c2}, \quad (7)$$

65 where Φ^{-1} is the azimuthal rotation matrix in the spin plane (around the spin axis in the coord-2 system) and Σ^{-1} is the transformation matrix to orient the z axis in the direction to the sensor package Pz direction. Again the matrices without inversion are used for the reconstruction of the model magnetic field in the calibration.

4. The field is further transformed into the spinning, non-orthogonal sensor-axis-aligned system (the coord-4 system in Fig. 1) by correcting for the elevation angles θ_1 (between the sensor-1 and the sensor-3 directions) and θ_2 (between the sensor-2 and the sensor-3 directions) and also for the azimuthal separation angle ϕ_{12} (between the sensor-1 and sensor-2 projected onto the plane normal to the sensor-3 direction) to obtain the magnetic field components B_1 , B_2 , and B_3 in the directions of the sensor axes including the gains and the offsets. The magnetic field vector in the coord-4 system is symbolically related to that in the coord-3 system as

$$B_{c4} = G^{-1} \Gamma^{-1} B_{c3} + O_s, \quad (8)$$

75 where Γ^{-1} is the transformation matrix using three angles (θ_1 , θ_2 , and ϕ_{12}), G^{-1} is the gain matrix, and O_s is the offset vector.

5. Finally, in the calibration procedure, the above transformations are inverted to estimate the ambient field from the sensor output. The estimated or reconstructed field is expressed the de-spin inertial coordinate system (the coord-5 system in Fig. 1) with the spin-plane primary component ($B_{X'}$), spin-plane residual component ($B_{Y'}$), and spin-axis component ($B_{Z'}$). The primed field expression in the coord-5 system ($B_{X'}$, $B_{Y'}$, and $B_{Z'}$) is identical to the model ambient field

B_X , B_Y , and B_Z in the coord-1 system if the calibration parameters are all accurately known. The model magnetic field is reconstructed from the sensor magnetic field as

$$B_{c5} = \Omega \Phi \Sigma \Gamma G (B_{c4} - O_s). \quad (9)$$

If the calibration parameters are all known, the reconstructed field B_{c5} restores the model field B_{c1} .

- 85 Note that the forward transformation is defined for the conversion of the sensor output (in the coord-4 system) into the magnetic field in the physically relevant system (the coord-1 system). In the error estimate study, the inverse transformation from the coord-1 system to the coord-4 system is more instructive in order to compare the calibrated magnetic field vector in the coord-5 system with the model ambient field in the coord-1 system.




Figure 1. Coordinate systems used in the magnetometer calibration error estimate.

- The relation between the sensor-output magnetic field $B_s = B_{c4}$ (introduced in the coord-4 system) and the model ambient field in the spinning frame B_{c2} (introduced in the coord-2 system, Eqs. 3–5) is expressed by a set of transformation matrices $G^{-1} \Gamma^{-1} \Sigma^{-1} \Phi^{-1}$ and an offset vector O_s as (Plaschke et al., 2019)

$$B_s = G^{-1} \Gamma^{-1} \Sigma^{-1} \Phi^{-1} B_{c2} + O_s. \quad (10)$$

- Here, the set of transformation matrices is composed of (1) the inverse rotation matrix around the spin axis Φ^{-1} by the rotation angle ϕ_a , (2) the inverse rotation matrix Σ^{-1} correcting for the tilt of spacecraft spin axis to the Pz direction (transforming the coord-2 system into the coord-3 system), (3) the inverse conversion matrix Γ^{-1} (transforming the coord-3 system into the

coord-4 system) and (4) the inverse gain matrix \mathbf{G}^{-1} . The sensor-output field is then corrected for the offset vector \mathbf{O}_s in the [sensor-axes directions](#). The matrices are constructed as follows (Plaschke et al., 2019).

$$\Phi^{-1} = \begin{pmatrix} 1 & \phi_a & 0 \\ -\phi_a & 1 & 0 \\ 0 & 0 & 1 \end{pmatrix} \quad (11)$$

$$100 \quad \Sigma^{-1} = \begin{pmatrix} 1 & 0 & \sigma_{Px} \\ 0 & 1 & \sigma_{Py} \\ -\sigma_{Px} & -\sigma_{Py} & 1 \end{pmatrix} \quad (12)$$

$$\Gamma^{-1} = \begin{pmatrix} 1 & 0 & -\delta\theta_1 \\ -\delta\phi_{12} & 1 & -\delta\theta_2 \\ 0 & 0 & 1 \end{pmatrix} \quad (13)$$

$$\mathbf{G}^{-1} = \begin{pmatrix} (gG_p)^{-1} & 0 & 0 \\ 0 & g^{-1}G_p & 0 \\ 0 & 0 & G_a^{-1} \end{pmatrix}. \quad (14)$$

The calibrated magnetic field vectors depend on the ambient magnetic field (B_p in the spin plane and B_a along the spin axis) and the following calibration parameters:

- 105
- gain ratio g between the two spin-plane sensors
 - absolute gains in the spin plane G_p and that in the spin axis direction G_a
 - offsets in the three sensor directions, O_1 , O_2 , and O_3
 - spin axis tilt angles σ_{Px} and σ_{Py} [to the angles in sensor package system](#) (σ_{Py} is the angle between the sensor-3 direction and the projection of the spin axis onto the sensor package Py–Pz plane; σ_{Px} is the angle of spin axis and the sensor package Py–Pz plane)
- 110
- deviation of elevation angles from 90° defined as $\delta\theta_1$ and $\delta\theta_2$, for the sensors 1 and 2, respectively
 - deviation of azimuthal angle from 90° defined as $\delta\phi_{12}$
 - rotation angle ϕ_a in the spin plane

Note that the orthogonality nearly holds such that the elevation and azimuthal angles exhibit only a small deviation from 90

115 degree,

$$\delta\theta_1 = \theta_1 - \frac{\pi}{2} \sim 0 \quad (15)$$

$$\delta\theta_2 = \theta_2 - \frac{\pi}{2} \sim 0 \quad (16)$$

$$\delta\phi_{12} = \phi_{12} - \frac{\pi}{2} \sim 0. \quad (17)$$

Also the tilt angles are small and close to zero,

$$120 \quad \sigma_{P_x} \sim 0 \quad (18)$$

$$\sigma_{P_y} \sim 0. \quad (19)$$

The relative gain and the two absolute gains are close to unity,

$$g \sim 1 \quad (20)$$

$$G_p \sim 1 \quad (21)$$

$$125 \quad G_a \sim 1. \quad (22)$$

The sensor output in the de-spun coordinate system (including the temperature dependence) is expressed up to the second lowest-order of the spin frequency as (Eqs. 24–26 in Plaschke et al., 2019):

$$130 \quad \begin{aligned} B_{X'} &= \frac{B_p(1+g^2)}{2gG_p} \\ &+ \left[O_1 + \frac{B_a(\sigma_{P_x} - \delta\theta_1)}{gG_p} \right] \cos\omega t \\ &- \left[O_2 + \frac{gB_a(\sigma_{P_y} - \delta\theta_2)}{G_p} \right] \sin\omega t \\ &+ \left[\frac{B_p(1-g^2)}{2gG_p} \right] \cos 2\omega t \\ &+ \frac{B_p}{2G_p} \left[g\phi_a - \frac{\phi_a}{g} + g\delta\phi_{12} \right] \sin 2\omega t \end{aligned} \quad (23)$$

$$135 \quad \begin{aligned} B_{Y'} &= -\frac{B_p}{2G_p} \left[\frac{1+g^2}{g} \phi_a + g\delta\phi_{12} \right] \\ &+ \left[O_2 + \frac{gB_a(\sigma_{P_y} - \delta\theta_2)}{G_p} \right] \cos\omega t \\ &+ \left[O_1 + \frac{B_a(\sigma_{P_x} - \delta\theta_1)}{gG_p} \right] \sin\omega t \\ &- \frac{B_p}{2G_p} \left[g\phi_a - \frac{\phi_a}{g} + g\delta\phi_{12} \right] \cos 2\omega t \\ &+ \left[\frac{B_p(1-g^2)}{2gG_p} \right] \sin 2\omega t \end{aligned} \quad (24)$$

$$140 \quad B_{Z'} = \frac{B_a}{G_a} + O_3 - \frac{B_p\sigma_{P_x}}{G_a} \cos\omega t + \frac{B_p\sigma_{P_y}}{G_a} \sin\omega t \quad (25)$$

Here, the magnetic field vector ($B_{X'}$, $B_{Y'}$, $B_{Z'}$) is represented in the coord-5 system and hence ideally reproduce the model magnetic field in the coord-1 system. That is, the z-component is in the direction of spacecraft spin axis, the x-component is in the spin plane. The y-component is also in the spin plane but should ideally not contain the ambient field. If the calibration parameters were all accurately known, the residual component ($B_{Y'}$) would be zero, and the ambient field reproduced or

145 reconstructed by the calibration has the spin-plane component ($B_{X'}$) and the spin-axis component ($B_{Z'}$). The directions of the three components are orthogonal if the calibration is accurate. Non-orthogonality may arise due to the uncertainties in the calibration parameters. The spacecraft spin frequency ω (as angular frequency) is assumed to be well known. t denotes the time in Eqs. (23)–(25). We also assume that the calibration parameters do not change over the time or along the spacecraft orbit.

2.1 Spin-plane primary component

150 The systematic error of magnetic field data is analytically derived by perturbing the calibration equations (Eqs. 23–25). The error in the X' component (spin-plane primary component) is denoted by $\Delta B_{X'}$. The spin-plane primary component is assumed to be aligned with the ambient field direction in the spin plane after calibration. On the assumption of the constant spin frequency ($\omega = \text{const.}$), the error $\Delta B_{X'}$ is derived by perturbing Eq. (23) as follows:

$$\begin{aligned}
 |\Delta B_{X'}| \leq & \max(\Delta O_1, \Delta O_2) + B_p \Delta \left(\frac{1}{2G_p} \left(\frac{1}{g} + g \right) \right) \\
 & + B_a \max \left(\Delta \left(\frac{1}{gG_p} |\sigma_{Px} - \delta\theta_1| \right), \Delta \left(\frac{g}{G_p} |\sigma_{Py} - \delta\theta_2| \right) \right) \\
 & + B_p \max \left(\Delta \left(\frac{1}{2G_p} \left| \frac{1}{g} - g \right| \right), \Delta \left(\frac{1}{2G_p} \left| \frac{1}{g} - g \right| \phi_a + g \delta\phi_{12} \right) \right)
 \end{aligned} \tag{26}$$

Here, the function $\max(x, y)$ returns the larger value from two variables, x and y , and is defined as

$$\max(x, y) = \frac{1}{2}(x + y + |x - y|) \tag{27}$$

The function $\max(x, y)$ takes the largest amplitude from an elliptically-shaped time series signal such as $x \cos(\omega t) + y \sin(\omega t)$.

160 After differential calculus (see Appendix), the expression of error $\Delta B_{X'}$ is arranged to that of calibration parameters (gains, offsets, and angles):

$$\begin{aligned}
 |\Delta B_{X'}| \leq & \max(\Delta O_1, \Delta O_2) \\
 & + B_p \frac{1}{2G_p^2} \left[\left(\frac{1}{g} + g \right) + \left| \frac{1}{g} - g \right| \max(1, \phi_a) \right] \Delta G_p \\
 & + B_a \frac{1}{G_p^2} \max \left(\frac{1}{g} |\sigma_{Px} - \delta\theta_1|, g |\sigma_{Py} - \delta\theta_2| \right) \Delta G_p \\
 & + B_p \left[\frac{1}{2G_p} \left| 1 - \frac{1}{g^2} \right| + \frac{1}{2G_p} \left(\frac{1}{g^2} + 1 \right) \max(1, \phi_a) + \delta\phi_{12} \right] \Delta g \\
 & + B_a \frac{1}{G_p} \max \left(\frac{1}{g^2} |\sigma_{Px} - \delta\theta_1|, |\sigma_{Py} - \delta\theta_2| \right) \Delta g \\
 & + B_a \frac{1}{G_p} \max \left(\frac{\Delta\sigma_{Px}}{g}, g \Delta\sigma_{Py} \right) \\
 & + B_a \frac{1}{G_p} \max \left(\frac{\Delta(\delta\theta_1)}{g}, g \Delta(\delta\theta_2) \right) \\
 & + B_p \frac{1}{2G_p} \left| g - \frac{1}{g} \right| \Delta\phi_a \\
 & + B_p g \Delta(\delta\phi_{12}).
 \end{aligned} \tag{28}$$

It is useful to introduce the following variables to simplify the notations:

$$\Delta O_{S1/2} = \max(\Delta O_1, \Delta O_2) \quad (29)$$

$$\Delta \sigma_{P_x/y} = \max(\Delta \sigma_{P_x}, \Delta \sigma_{P_y}) \quad (30)$$

$$\Delta(\theta_{S1/2}) = \max(\Delta(\delta\theta_1), \Delta(\delta\theta_2)) \quad (31)$$

- 175 If the gains (both absolute and relative ones) are close to unity ($g \simeq 1$, $G_p \simeq 1$) and the misalignments are small ($\sigma_{P_x} \ll 1$ rad, $\sigma_{P_y} \ll 1$ rad, $\delta\theta_1 \ll 1$ rad, $\delta\theta_2 \ll 1$ rad, $\delta\phi_{12} \ll 1$ rad), Eq. (28) is simplified with the leading terms:

$$\begin{aligned} |\Delta B_{X'}| &\leq \Delta O_{S1/2} \\ &+ B_p (\Delta G_p + \max(1, \phi_a) \Delta g + \Delta(\delta\phi_{12})) \\ &+ B_a (\Delta \sigma_{P_x/y} + \Delta(\delta\theta_{S1/2})). \end{aligned} \quad (32)$$

- 180 We assume $\max(1, \phi_a) = 1$ (which is realized when $\phi_a \leq 1$ holds), then Eq. (32) is further simplified into a more practical form:

$$\begin{aligned} |\Delta B_{X'}| &\leq \Delta O_{S1/2} \\ &+ B_p (\Delta G_p + \Delta g + \Delta(\delta\phi_{12})) \\ &+ B_a (\Delta \sigma_{P_x/y} + \Delta(\delta\theta_{S1/2})). \end{aligned} \quad (33)$$

185 2.2 Spin-plane residual component

Derivation of the error in the Y' component (which residual to the primary component after determination or reconstruction of the ambient field in the spin plane) nearly follows that in the X' component. Note that the Y' component has only a tiny amount of the ambient field because of its residual character. The Y' component vanishes if the calibration is properly and accurately done. After derivative calculations (see Appendix), the error of the residual component is estimated as

$$\begin{aligned} 190 \quad |\Delta B_{Y'}| &\leq \max(\Delta O_1, \Delta O_2) \\ &+ B_p \Delta \left(\frac{1}{2G_p} \left(\left(\frac{1}{g} + g \right) \phi_a + g \delta\phi_{12} \right) \right) \\ &+ B_a \max \left(\Delta \left(\frac{1}{gG_p} |\sigma_{P_x} - \delta\theta_1| \right), \Delta \left(\frac{g}{G_p} |\sigma_{P_y} - \delta\theta_2| \right) \right) \\ &+ B_p \max \left(\Delta \left(\frac{1}{2G_p} \left| \frac{1}{g} - g \right| \right), \Delta \left(\frac{1}{2G_p} \left| \frac{1}{g} - g \right| \phi_a + g \delta\phi_{12} \right) \right) \end{aligned} \quad (34)$$

Equation (34) is sorted to the errors of calibration parameters as:

$$\begin{aligned}
195 \quad |\Delta B_{Y'}| &\leq \max(\Delta O_1, \Delta O_2) \\
&+ B_p \frac{1}{2G_p^2} \left[\left(\frac{1}{g} + g \right) \phi_a + \left| \frac{1}{g} - g \right| \max(1, \phi_a) \right] \Delta G_p \\
&+ B_a \frac{1}{G_p^2} \max \left(\frac{1}{g} |\sigma_{Px} - \delta\theta_1|, g |\sigma_{Py} - \delta\theta_2| \right) \Delta G_p \\
&+ B_p \left[\frac{1}{2G_p} \left| 1 - \frac{1}{g^2} \right| \phi_a + \frac{1}{2G_p} \left(\frac{1}{g^2} + 1 \right) \max(1, \phi_a) + 2 \delta\phi_{12} \right] \Delta g \\
&+ B_a \frac{1}{G_p} \max \left(\frac{1}{g^2} |\sigma_{Px} - \delta\theta_1|, |\sigma_{Py} - \delta\theta_2| \right) \Delta g \\
200 \quad &+ B_a \frac{1}{G_p} \max \left(\frac{\Delta\sigma_{Px}}{g}, g \Delta\sigma_{Py} \right) \\
&+ B_a \frac{1}{G_p} \max \left(\frac{\Delta(\delta\theta_1)}{g}, g \Delta(\delta\theta_2) \right) \\
&+ B_p \frac{1}{2G_p} \left[\left(\frac{1}{g} + g \right) + \left| g - \frac{1}{g} \right| \right] \Delta\phi_a \\
&+ B_p 2g \Delta(\delta\phi_{12}). \tag{35}
\end{aligned}$$

Again, as done in the calculation of the X' component, we take the leading terms (the first order terms) and obtain a simplified
205 expression of the error of residual component as:

$$\begin{aligned}
|\Delta B_{Y'}| &\leq \Delta O_{S1/2} \\
&+ B_p (\Delta G_p + \Delta g + 2 \Delta(\delta\phi_{12}) + \Delta\phi_a) \\
&+ B_a (\Delta\sigma_{Px/y} + \Delta(\delta\theta_{S1/2})). \tag{36}
\end{aligned}$$

The differences from $\Delta B_{X'}$ (Eq. 33) are $2\Delta(\delta\phi_{12})$ and $\Delta\phi_a$ in the second term in Eq. (36). The appearance of $\Delta\phi_a$ means that
210 the uncertainty of the magnetometer boom extension angle (the spin-plane rotation angle) causes a finite residual component,
that is, the spin-plane ambient field is erroneously projected to yield the residual component Y' by an angle of $\Delta\phi_a$. The effect
of $\Delta\phi_a$ on the spin-plane primary field component is of the second order, while that on the residual component is of the first
order. According to our estimate of the calibration parameter errors (Tab. 1), the first-order errors are in the range between
 10^{-2} and 10^{-4} and the **second order** errors (due to the couplings of calibration errors with the other small parameters) are in
215 the range between 10^{-5} and 10^{-8} .

2.3 Spin-axis component

The error of spin-axis component is derived from Eq. (25) in a straightforward fashion:

$$|\Delta B_{Z'}| \leq \Delta O_3 + B_a \Delta \left(\frac{1}{G_a} \right) + B_p \max \left(\Delta \left(\frac{\sigma_{Px}}{G_a} \right), \Delta \left(\frac{\sigma_{Py}}{G_a} \right) \right) \quad (37)$$

220

$$\begin{aligned} &\leq \Delta O_3 \\ &\quad + B_a \frac{1}{G_a^2} \Delta G_a \\ &\quad + B_p \frac{1}{G_a^2} \max(\sigma_{Px}, \sigma_{Py}) \Delta G_a \\ &\quad + B_p \frac{1}{G_a} \max(\Delta \sigma_{Px}, \Delta \sigma_{Py}) \end{aligned} \quad (38)$$

For a nearly unit gain in the axial direction ($G_a \simeq 1$) and small misalignments ($\sigma_{Px} \ll 1$, $\sigma_{Py} \ll 1$), the expression of error estimate is simplified into:

$$|\Delta B_{Z'}| \leq \Delta O_3 + B_a \Delta G_a + B_p \Delta \sigma_{Px/y}. \quad (39)$$

Equation (39) indicates that an error occurs in the spin-axis direction (1) when the offset ΔO_3 is present, (2) when the axial (absolute) gain G_a has an uncertainty, or (3) when the spin axis angle relative to the sensor Z direction has an uncertainty (which introduces a mixing or projection of the spin-plane component by the spin-axis component).

3 Estimate of calibration parameter errors

230 Nominal errors (as upper limits) of calibration parameters are summarized in Tab. 1 as lessons from Earth-orbiting spinning spacecraft Cluster (Escoubet et al., 2001; Balogh et al., 2001), THEMIS (Angelopoulos, 2008; Auster et al., 2008), and MMS (Burch et al., 2016; Russell et al., 2016). The spin-plane-related calibration parameters are assessed in detail by Plaschke et al. (2019). The accuracy studies on the spin-axis offset are presented by Alconcel et al. (2014), Frühauff et al. (2017), Plaschke (2019), and Schmid et al. (2020). In the following, we review the uncertainties of calibration parameters.

235 3.1 Offset error

The offsets in the spin plane (O_1 and O_2) are determined by the in-flight calibration. The error of spin-plane offsets on in-flight calibration is, after Plaschke et al. (2019), minimized down to the sum of (1) spin-plane component of natural fluctuation at the spin frequency (denoted by F_p), (2) projection of spin-axis ambient field by an error of spin-axis angle $B_a \Delta \sigma_{Px/y}$, and (3) projection of spin-axis ambient field by an error sensor elevation angle $B_a \Delta(\delta\theta_{S1/2})$:

$$\Delta O_{S1/2} \simeq F_p + B_a \Delta \sigma_{Px/y} + B_a \Delta(\delta\theta_{S1/2}). \quad (40)$$

The lesson from the in-flight calibration for the THEMIS magnetometer data indicates that an offset value of about 0.1 nT or better (i.e., smaller) can be reached using spacecraft spin (Plaschke et al., 2019).

Table 1. Nominal errors of calibration parameters. Five lines from the top (spin axis angles, gain ratio, azimuthal angle, spin-plane offsets, and elevation angles) represent the in-flight calibration for THEMIS (Plaschke et al., 2019). Nominal error of spin-axis offset may vary between 0.2 nT in the solar wind (Plaschke, 2019) and 1 nT in the magnetosphere from temperature drift studies by Alconcel et al. (2014) and Frühauff et al. (2017). Absolute gains in the spin plane and along the spin axis are taken from the ground calibration experience. Spin-plane rotation angle is taken from the magnetometer boom design for BepiColombo Mio.

Parameter	Symbol	Error upper limit
Spin axis angle (x or y directions)	$\Delta\sigma_{P_{x/y}}$	10^{-4} rad
Gain ratio	Δg	10^{-4}
Azimuthal angle	$\Delta(\delta\phi_{12})$	10^{-4} rad
Spin-plane offset S1 or S2	$\Delta O_{S1/2}$	0.1 nT
Elevation angle S1 or S2	$\Delta(\delta\theta_{S1/2})$	10^{-3} rad
Spin-axis offset S3 (solar wind)	$\Delta O_3^{(sw)}$	0.2 nT
Spin-axis offset S3 (magnetosphere)	$\Delta O_3^{(ms)}$	1 nT
Spin-plane absolute gain	ΔG_p	10^{-3}
Spin-axis absolute gain	ΔG_a	10^{-3}
Spin-plane rotation angle	$\Delta\phi_a$	10^{-2} rad

The offset in the spin-axis direction cannot be determined from the spacecraft spin, but needs to be determined in different ways, for example, using additional measurements such as absolute magnetic field magnitude (Nakamura et al., 2014; Plaschke et al., 2014) or using plasma physical properties such as the nearly-incompressible fluctuation nature in the solar wind (Hedgecock, 1975; Leinweber et al., 2008), the highly-compressible fluctuation nature in which the fluctuations are nearly aligned with the ambient field (Plaschke and Narita, 2016; Plaschke et al., 2017), or the magnetic null environment in diamagnetic cavities around comets (Goetz et al., 2016a, b). The uncertainty in the spin-axis offset can empirically be minimized to 0.2 nT when using the solar wind fluctuations (Plaschke, 2019) and the mirror-mode fluctuations (Plaschke and Narita, 2016; Frühauff et al., 2017). The accuracy of spin-axis offset determination can be improved when a larger amount of data is available. An accuracy of 0.5 nT or 1.0 nT is considered as representative using the mirror-mode fluctuations (Schmid et al., 2020). It is also worth noting that the offset drift is up to 1 nT per year as lessons from Cluster (Alconcel et al., 2014) and THEMIS (Frühauff et al., 2017), which may be used as a nominal value of spin-axis offset error when the spacecraft stays in the magnetosphere and the in-situ offset determination using solar wind or mirror-mode fluctuations is not possible.

3.2 Gain error

The error of gain ratio in the spin plane is minimized to the natural fluctuation amplitude at the second harmonic of spin frequency in the spin plane (denoted by F_{2p}) relative to the spin-plane ambient field B_p (Plaschke et al., 2019):

$$\Delta g \simeq \frac{F_{2p}}{B_p} \quad (41)$$

The gain ratio can be determined to a reasonably accurate level using the spacecraft spin, down to an uncertainty of about 10^{-4} (Plaschke et al., 2019). It is true that the gain ratio in the spin plane g is related to the sensitivity measurements during the ground calibration through:

$$g^2 = \frac{S_x}{S_y}, \quad (42)$$

where S_x and S_y are the sensitivity (absolute gain) of the two spin-plane sensors, but the gain ratio obtained from the in-flight calibration is sufficiently accurate ($\Delta g \simeq 10^{-4}$) in practical applications.

265 3.3 Sensor axis non-orthogonality

Sensor axis non-orthogonality includes errors of the elevation angles $\Delta(\delta\theta_1)$ and $\Delta(\delta\theta_2)$ and azimuthal angles between S1 and S2 $\Delta(\delta\phi_{12})$. The error of elevation angles $\Delta(\delta\theta_1)$ and $\Delta(\delta\theta_2)$ is, after Plaschke et al. (2019), minimized to the sum of (1) natural frequency at the spin frequency relative to the ambient spin-axial field, (2) offset error relative to the ambient spin-axial field, and (3) uncertainty of the spin-axis angle as

$$270 \quad \Delta(\delta\theta_{S1/2}) \simeq \frac{F_p}{B_a} + \frac{\Delta O_{S1/2}}{B_a} + \Delta\sigma_{P_{x/y}}. \quad (43)$$

The elevation angles $\Delta(\delta\theta_1)$ and $\Delta(\delta\theta_2)$ are the angles between the sensors S1 and S3, and that between S2 and S3, respectively. The angle uncertainties $\Delta(\delta\theta_1)$ and $\Delta(\delta\theta_2)$ can be obtained both from the ground calibration and from the in-flight calibration. Errors of the elevation angles are about 10^{-3} in the in-flight calibration (Plaschke et al., 2019).

The azimuthal angle deviation $\delta\phi_{12}$ is also related to the ground-calibrated sensor angles ξ_{12} , ξ_{13} , and ξ_{23} . It is straightforward to show, by using the trigonometric relations, that the relation is

$$\sin(\delta\phi_{12}) = \sin(\delta\xi_{12}) + \sin(\delta\xi_{13})\sin(\delta\xi_{23}) \quad (44)$$

For smaller deviation angles of ϕ_{S12} , ξ_{12} , ξ_{13} , and ξ_{23} (i.e., if the sensors are nearly orthogonal to one another), the relation is simplified into

$$\Delta(\delta\phi_{S12}) \simeq \Delta(\delta\xi_{12}). \quad (45)$$

280 The azimuthal angle $\delta\phi_{S12}$ can thus be obtained both from the ground calibration and from the in-flight calibration, and its uncertainty can be sufficiently minimized down to about 10^{-4} rad in the in-flight calibration (Plaschke et al., 2019).

3.4 Misalignment to the spacecraft reference direction

Angular deviation of the the spin axis from the normal direction of the sensor x - y plane is characterized by two angles, σ_{P_x} and σ_{P_y} . The error of misalignment angles σ_{P_x} and σ_{P_y} is estimated as the ratio of the spin-axis natural fluctuation amplitude at the spin frequency to the spin-plane ambient field,

$$285 \quad \sigma_{P_{x/y}} \simeq \frac{F_a}{B_p}, \quad (46)$$

and the value of $\sigma_{P_{x/y}}$ is empirically about 10^{-4} rad (Plaschke et al., 2019). The angles σ_{P_x} and σ_{P_y} need the determination or knowledge of spacecraft spin axis, and cannot usually be evaluated during the ground calibration of the sensors.

The remaining angle is the rotation angle in the spin plane. The rotation angle can be determined in flight using Earth's
 290 magnetic field model in the case of Earth-orbiting spacecraft, and the method works better in a high-field environment. For
 example, the rotation angle is determined to an **accuracy of 0.5°** or better when using the magnetic field data around the perigee
 with a field magnitude of about 8000 nT. In-flight determination of the rotation angle is meaningful when the accuracy in the
 in-flight method is better than the knowledge from the boom design with ground verification. We take the case of BepiColombo
 Mio magnetometer because the magnetometer boom extension direction is known to be within an uncertainty of 0.5° (which
 295 gives $\Delta\phi_a = 8.7 \times 10^{-3}$ rad $\simeq 10^{-2}$ rad) from the spacecraft design and ground verification. As we will see in the next section,
 the uncertainty of rotation angle in the spin plane plays an important role in the final error estimate in a high-field environment.

4 Combined errors of calibrated magnetometer data

The individual error sources are combined using the first-order expressions (Eqs. 33, 36, and 39) to evaluate the error of
 calibrated magnetometer data for the nominal parameters (Tab. 1). Here, the errors represent the upper limits of the three
 300 magnetic field data in three directions (spin-plane primary, spin-plane residual, and spin-axis components). **For a practical
 purpose, the combined errors in Eqs. (33), (36), and (39) are reformulated in an approximate form using the values given in
 Tab. 1:**

$$|\Delta B_{x'}| \leq 0.1 \text{ [nT]} + (B_p + B_a) \times 10^{-2} \quad (47)$$

$$|\Delta B_{y'}| \leq 0.1 \text{ [nT]} + (10B_p + B_a) \times 10^{-3} \quad (48)$$

$$305 \quad |\Delta B_{z'}| \leq 0.2 \text{ [nT]} + (B_p + B_a) \times 10^{-3}. \quad (49)$$

The combined errors are graphically displayed in Fig. 2 as a function of the ambient magnetic field in the spin-axis direction
 (0° , data curves in black) and spin-plane direction (90° , data curves in gray). Equations (33), (36), and (39) and Fig. 2 indicate
 that the calibration error has two distinct domains: (1) the offset dominant domain in a low-field, up to an ambient field of about
 1 nT when the field is along the spin axis (curves in black in Fig. 2), and up to 10 nT when the field is in the spin plane (curves
 310 in gray in Fig. 2), and (2) the ambient field-dependent domain in a high field (above 1 or 10 nT). In the low-field case, the offset
 dominates the magnetometer data error and the offset value is expected in the range between 0.1 to 1 nT. In the high-field case,
 the error grows linearly with the ambient field, and the relative error is expected between 1% (which comes from $\Delta\phi_a$) and
 0.1% (which comes from the absolute gain error and the elevation angle error).

The error depends on the angle between the ambient field and the spacecraft spin axis. The gain errors, azimuthal angle error,
 315 and boom misalignment are coupled to the spin-plane ambient field in the spin-plane components (Eqs. 33 and 36). The spin
 axis misalignment and elevation angle errors are coupled to the spin-axis field. The axial gain and the spin axis misalignment
 are coupled to the spin-axis and spin-plane ambient field, respectively, in the expression of spin-axis component (Eq. 39).




Figure 2. Error of in-flight calibrated magnetometer data for an error of magnetometer boom angle $\delta\phi_a \leq 0.5^\circ \sim 10^{-2}$ rad (the case for the BepiColombo Mio magnetometer) Curves in black and in gray represent for the axial ambient magnetic field (0° to the spin axis) and the spin-plane ambient field (90°), respectively.

The residual component has the largest uncertainty in Fig. 2, which comes from the uncertainty of spin-plane rotation angle $\Delta\phi_a$. For the reference purpose, Figure 3 exhibits the combined error estimate for the error of azimuthal angle smaller than that for Fig. 2 by an order of magnitude, $\delta\phi_a \sim 10^{-3}$ rad. In that case, the angle errors in the calibration parameters fall onto the nearly same order (between 10^{-4} rad and 10^{-3} rad). The final error is then below 1 nT (up to an ambient field of 300 nT) even when the ambient field is along the spin axis.

The graphical representation of the error estimates is extended to an ambient field of up to 10,000 nT, and is plotted again for different values of rotation angle ($\Delta\phi_a = 10^{-2}$ rad and $\Delta\phi_a = 10^{-3}$ rad) in Figs. 4 and 5, respectively,

325 5 Conclusions

Fluxgate magnetometers are widely used in a wide range of spacecraft missions for the studies of Earth's and planetary magnetospheres, solar system bodies, and heliosphere. Magnetometer and the associated calibration process are necessarily accompanied by uncertainties that arise from various error sources. We conclude the error estimate on magnetometer in-flight calibration as follows.

- 330 1. Errors appear both as absolute ones (which are the offsets) and as relative ones (angle errors, gain errors). First-order expressions (Eqs. 33–39) (also graphically displayed in Figs. 2–5) are of practical use, and show that the offset errors




Figure 3. The same plot style as Fig. 2 but for the improved error of magnetometer boom angle $\delta\phi_a \leq 0.05^\circ \sim 10^{-3}$ rad.




Figure 4. The same plot style as Fig. 2 but for an extended ambient field up to 10,000 nT.

dominate in a low ambient field (typically below 10 nT) while the relative errors (proportional to the ambient field) dominate in a high ambient field.




Figure 5. The same plot style as Fig. 3 but for an extended ambient field up to 10,000 nT.

2. The largest uncertainty sources are (1) the spin-axis offset error and (2) the spin-plane rotation angle error. The offset error appears as the dominant error in the low-field environment. The spin-plane rotation angle error plays a major role in a high-field environment, particularly when the ambient field is aligned in the spin plane.

The uncertainties are obtained by perturbing the calibration parameters proposed by Plaschke et al. (2019). When simplified into the first-order expression, the magnetometer data errors primarily represent the offset errors as constant and the errors of gains and angles as relative error to the ambient field. Our derivation shows how the uncertainty sources combine through the calibration process both linearly (which is dominant) and non-linearly through coupling of calibration parameter errors (which is of only secondary importance when the errors of calibration parameters are small). The error formulas are presented with analytical expressions (Eqs. 33, 36, and 39), and are expected to serve as a useful tool in various applications, for example, to further minimize the final error in designing a magnetometer with a boom and verifying the error thoroughly in the ground calibration (particularly the spin-plane rotation angle) and to report the error of scientific studies which are based on magnetometer data.

It should be noted that the calibration parameters are treated as time independent in our study. In reality, however, the calibration parameters (such as offsets and gains) depend on the temperature and can evolve along the orbit. Time-dependent picture of the calibration parameters needs an extensive in-flight calibration experience.

The errors associated with the uncertainties in calibration parameters are studied in this paper. In a low-field environment such as in interplanetary space the sensor nonlinearity (which originates in the nonlinearity of gain) is usually considered negligible. In a low Earth orbit the situation may be different. Modern sensors which are often double wound, and even triple

would have excellent linearity (typically to an accuracy of about 10^{-4} per axis), but this is not always the case. The MAGSAT single-wound sensor (Acuña, 1980; Langel et al., 1982), for example, suffered from about 1% nonlinearity, and the same sensor design was used more recently on MESSENGER (Solomon et al., 2007; Anderson et al., 2007). With present thinking about
 355 the possibility of deploying large fleets of small magnetometer cubesats with just as small sensors one might ask whether nonlinearity issues can rise again.

Acknowledgements. This work was financially supported by the Austrian Space Applications Programme (ASAP) at the Austrian Research Promotion Agency under contract 865967. YN also acknowledges financial support by the Japan Society for the Promotion of Science, Invitational Fellowship for Research in Japan (long term) under grant FY2019 L19527. YN also thanks the research and administration staff
 360 members in Hoshino laboratory group at the University of Tokyo for discussions, supports, and organizations during the fellowship program and the University of Tokyo Mejirodai International Village (MIV) for the arrangement and hospitality during the pleasant and productive stay in Tokyo.

Appendix A: Derivatives

Detailed derivative calculations in section 2 are presented here.

$$365 \quad \Delta \left(\frac{1}{2G_p} \left(\frac{1}{g} + g \right) \right) \leq \frac{1}{2G_p^2} \left(\frac{1}{g} + g \right) \Delta G_p + \frac{1}{2G_p} \left| 1 - \frac{1}{g} \right| \Delta g \quad (\text{A1})$$

$$\begin{aligned} \Delta \left(\frac{1}{gG_p} (\sigma_{P_x} - \delta\theta_1) \right) &= \Delta \left(\frac{1}{g} \right) \frac{1}{G_p} (\sigma_{P_x} - \delta\theta_1) \\ &\quad + \frac{1}{g} \Delta \left(\frac{1}{G_p} \right) (\sigma_{P_y} - \delta\theta_1) \\ &\quad + \frac{1}{gG_p} \Delta (\sigma_{P_x} - \delta\theta_1) \end{aligned} \quad (\text{A2})$$

$$\begin{aligned} &\leq \frac{1}{g^2G_p} |\sigma_{P_x} - \delta\theta_1| \Delta g \\ 370 \quad &\quad + \frac{1}{gG_p^2} |\sigma_{P_x} - \delta\theta_1| \Delta G_p \\ &\quad + \frac{1}{gG_p} \Delta \sigma_{P_x} + \frac{1}{gG_p} \Delta (\delta\theta_1) \end{aligned} \quad (\text{A3})$$

$$\begin{aligned}
\Delta\left(\frac{g}{G_p}(\sigma_{Py} - \delta\theta_2)\right) &= (\Delta g)\frac{1}{G_p}(\sigma_{Py} - \delta\theta_2) \\
&\quad + g\Delta\left(\frac{1}{G_p}\right)(\sigma_{Py} - \delta\theta_2) \\
&\quad + \frac{g}{G_p}\Delta(\sigma_{Py} - \delta\theta_2)
\end{aligned} \tag{A4}$$

375

$$\begin{aligned}
&\leq \frac{1}{G_p}|\sigma_{Py} - \delta\theta_2| \Delta g \\
&\quad + \frac{g}{G_p^2}|\sigma_{Py} - \delta\theta_2| \Delta G_p \\
&\quad + \frac{g}{G_p}\Delta\sigma_{Py} + \frac{g}{G_p}\Delta(\delta\theta_2 <)
\end{aligned} \tag{A5}$$

$$\Delta\left(\frac{1}{2G_p}\left(\frac{1}{g} - g\right)\right) = \Delta\left(\frac{1}{2G_p}\right)\left(\frac{1}{g} - g\right) + \frac{1}{2G_p}\Delta\left(\frac{1}{g} - g\right) \tag{A6}$$

$$\leq \frac{1}{2G_p^2}\left|\frac{1}{g} - g\right| \Delta G_p + \frac{1}{2G_p}\left(\frac{1}{g^2} + 1\right) \Delta g \tag{A7}$$

$$\begin{aligned}
380 \quad \Delta\left(\frac{1}{2G_p}\left(g - \frac{1}{g}\right)\phi_a + g\delta\phi_{S12}\right) &= \Delta\left(\frac{1}{2G_p}\right)\left(g - \frac{1}{g}\right)\phi_a + \frac{1}{2G_p}\Delta\left(g - \frac{1}{g}\right)\phi_a \\
&\quad + \frac{1}{2G_p}\left(g - \frac{1}{g}\right)\Delta\phi_a + \delta\phi_{S12}\Delta g + g\Delta(\delta\phi_{S12})
\end{aligned} \tag{A8}$$

$$\begin{aligned}
&\leq \frac{1}{2G_p^2}\left|g - \frac{1}{g}\right|\phi_a \Delta G_p + \frac{1}{2G_p}\phi_a\left(1 + \frac{1}{g^2}\right)\Delta g \\
&\quad + \frac{1}{2G_p}\left|g - \frac{1}{g}\right|\Delta\phi_a + \delta\phi_{S12}\Delta g + g\Delta(\delta\phi_{S12})
\end{aligned} \tag{A9}$$

$$\begin{aligned}
\Delta\left(\frac{1}{2G_p}\left(\frac{1}{g} + g\right)\phi_a + g\delta\phi_{S12}\right) &= \Delta\left(\frac{1}{2G_p}\right)\left(\frac{1}{g} + g\right)\phi_a + \frac{1}{2G_p}\phi_a\Delta\left(\frac{1}{g} + g\right) \\
&\quad + \frac{1}{2G_p}\left(\frac{1}{g} + g\right)\Delta\phi_a + \delta\phi_{S12}\Delta g + g\Delta(\delta\phi_{S12})
\end{aligned} \tag{A10}$$

385

$$\begin{aligned}
&\leq \frac{1}{2G_p^2}\left(\frac{1}{g} + g\right)\phi_a \Delta G_p + \frac{1}{2G_p}\left(1 - \frac{1}{g^2}\right)\Delta g \\
&\quad + \frac{1}{2G_p}\left(\frac{1}{g} + g\right)\Delta\phi_a + \delta\phi_{S12}\Delta g + g\Delta(\delta\phi_{S12})
\end{aligned} \tag{A11}$$

References

- Acuña, M. H.: The MAGSAT precision vector magnetometer, *Johns Hopkins APL Technical Digest*, 1, 3, 210–213, 1980.
- 390 Acuña, M. H.: Space-based magnetometers, *Rev. Sci. Instrum.*, 73, 3717, 2002. <https://doi.org/10.1063/1.1510570>
- Alconcel, L. N. S., Fox, P., Brown, P., Oddy, T. M., Lucek, E. L., and Carr, C. M.: An initial investigation of the long-term trends in the fluxgate magnetometer (FGM) calibration parameters on the four Cluster spacecraft, *Geosci. Instrum. Method. Data Syst.*, 3, 95–109, 2014. <https://doi.org/10.5194/gi-3-95-2014>
- Anderson, B. J., Acuña, M. H., Lohr, D. A., Scheifele, J., Raval, A., Korth, H., and Slavin, J. A.: The magnetometer instrument on MES-
395 SENER, *Space Sci. Rev.*, 131, 417–450, 2007. <https://doi.org/10.1007/s11214-007-9246-7>
- Angelopoulos, V.: The THEMIS mission, *Space Sci. Rev.*, 141, 5, 2008. <https://doi.org/10.1007/s11214-008-9336-1>
- Auster, H. U., Glassmeier, K. H., Magnes, W., Aydogar, O., Baumjohann, W., Constantinescu, D., Fischer, D., Fornacon, K. H., Georgescu, E., Harvey, P., Hillenmaier, O., Kroth, R., Ludlam, M., Narita, Y., Nakamura, R. Okrafka, K., Plaschke, F., Richter, I., Schwarzl, H., Stoll, B., Valavanoglou, A. and Wiedemann, M.: The THEMIS Fluxgate Magnetometer, *Space Sci. Rev.*, 141, 235–264, 2008.
400 <https://doi.org/10.1007/s11214-008-9365-9>
- Balogh, A., Carr, C. M., Acuña, M. H., Dunlop, M. W., Beek, T. J., Brown, P., Fornacon, K.-H., Georgescu, E., Glassmeier, K.-H., Harris, J., Musmann, G., Oddy, T., and Schwingenschuh, K.: The Cluster Magnetic Field Investigation: overview of in-flight performance and initial results, *Ann. Geophys.*, 19, 1207–1217, 2001. <https://doi.org/10.5194/angeo-19-1207-2001>
- Burch, J. L., Moore, T. E., Torbert, R. B., and Giles, B. L.: Magnetospheric Multiscale overview and science objectives, *Space Sci. Rev.*, 199,
405 5–21, 2016. <https://doi.org/10.1007/s11214-015-0164-9>
- Escoubet, C. P., Fehringer, M., and Goldstein, M.: Introduction – The Cluster mission, *Ann. Geophys.*, 19, 1197–1200, 2001. <https://10.5194/angeo-19-1197-2001>
- Frühauff, D., Plaschke, F., and Glassmeier, K.-H.: Spin axis offset calibration on THEMIS using mirror modes, *Ann. Geophys.*, 35, 117–121, 2017. <https://doi.org/10.5194/angeo-35-117-2017>
- 410 Goetz, C., Koenders, C., Hansen, K. C., Burch, J., Carr, C., Eriksson, A., Frühauff, D., Güttler, C., Henri, P., Nilsson, H., Richter, I., Rubin, M., Sierks, H., Tsurutani, B., Volwerk, M., and Glassmeier, K. H.: Structure and evolution of the diamagnetic cavity at comet 67P/Churyumov-Gerasimenko, *Mon. Not. R. Astron. Soc.*, 462, S459–S467, 2016a. <https://doi.org/10.1093/mnras/stw3148>
- Goetz, C., Koenders, C., Richter, I., Altwegg, K., Burch, J., Carr, C., Cupido, E., Eriksson, A., Güttler, C., Henri, P., Mokashi, P., Nemeth, Z., Nilsson, H., Rubin, M., Sierks, H., Tsurutani, B., Vallat, C., Volwerk, M., and Glassmeier, K.-H.: First detection of a diamagnetic cavity
415 at comet 67P/Churyumov-Gerasimenko, *Astron. Astrophys.*, 588, A24, 2016b. <https://doi.org/10.1051/0004-6361/201527728>
- Hedgecock, P. C.: A correlation technique for magnetometer zero level determination, *Space Sci. Instrum.*, 1, 83–90, 1975.
- Kepko, E. L., Khurana, K. K., Kivelson, M. G., Elphic, R. C., and Russell, C. T.: Accurate determination of magnetic field gradients from four point vector measurements. I. Use of natural constraints on vector data obtained from a single spinning spacecraft, *IEEE T. Magnet.*, 32, 377–385, 1996. <https://doi.org/10.1109/20.486522>
- 420 Langel R., Ousley G., Berbert J., Murphy J., and Settle M.: The MAGSAT Mission. *Geophys. Res. Lett.*, 9, 243–245, 1982. <https://doi.org/10.1029/GL009i004p00243>
- Leinweber, H. K., Russell, C. T., Torkar, K., Zhang, T. L., and Angelopoulos, V.: An advanced approach to finding magnetometer zero levels in the interplanetary magnetic field, *Meas. Sci. Technol.*, 19, 055104, 2008. <https://doi.org/10.1088/0957-0233/19/5/055104>

- Nakamura, R., Plaschke, F., Teubenbacher, R., Giner, L., Baumjohann, W., Magnes, W., Steller, M., Torbert, R. B., Vaith, H., Chutter, M., Fornaçon, K.-H., Glassmeier, K.-H., and Carr, C.: Interinstrument calibration using magnetic field data from the fluxgate magnetometer (FGM) and electron drift instrument (EDI) onboard Cluster, *Geosci. Instrum., Method. Data Syst.*, 3, 1–11, 2014. <https://doi.org/10.5194/gi-3-1-2014>
- 425 Plaschke, F., Nakamura, R., Leinweber, H. K., Chutter, M., Vaith, H., Baumjohann, W., Steller, M., and Magnes, W.: Fluxgate magnetometer spin axis offset calibration using the electron drift instrument, *Meas. Sci. Technol.*, 25, 105008, 2014. [https://doi.org/10.1088/0957-](https://doi.org/10.1088/0957-0233/25/10/105008)
- 430 0233/25/10/105008
- Plaschke, F. and Narita, Y.: On determining fluxgate magnetometer spin axis offsets from mirror mode observations, *Ann. Geophys.*, 34, 759–766, 2016. <https://doi.org/10.5194/angeo-34-759-2016>
- Plaschke, F., Goetz, C., Volwerk, M., Richter, I., Frühauff, D., Narita, Y., Glassmeier, K.-H., and Dougherty, M. K.: Fluxgate magnetometer offset vector determination by the 3D mirror mode method, *Mon. Notic. Roy. Astron. Soc.*, 469, S675–S684, 2017. <https://doi.org/10.1093/mnras/stx2532>
- 435 Plaschke, F., Auster, H.-U., Fischer, D., Fornaçon, K.-H., Magnes, W., Richter, I., Constantinescu, D., and Narita, Y.: Advanced calibration of magnetometers on spin-stabilized spacecraft based on parameter decoupling, *Geosci. Instrum. Method. Data Syst.*, 8, 63–76, 2019. <https://doi.org/10.5194/gi-8-63-2019>
- Plaschke, F.: How many solar wind data are sufficient for accurate fluxgate magnetometer offset determinations? *Geosci. Instrum. Method. Data Syst.*, 8, 285–291, 2019. <https://doi.org/10.5194/gi-8-285-2019>
- 440 Russell, C. T., Anderson, B. J., Baumjohann, W., Bromund, K. R., Dearborn, D., Fischer, D., Le, G., Leinweber, H. K., Leneman, D., Magnes, W., Means, J. D., Moldwin, M. B., Nakamura, R., Pierce, D., Plaschke, F., Rowe, K. M., Slavin, J. A., Strangeway, R. J., Torbert, R., Hagen, C., Jernej, I., Valavanoglou, A., and Richter, I.: The Magnetospheric Multiscale Magnetometers, *Space Sci. Rev.*, 199, 189–256, 2016. <https://doi.org/10.1007/s11214-014-0057-3>
- 445 Schmid, D., Plaschke, F., Narita, Y., Heyner, D., Mieth, J. Z. D., Anderson, B. J., Volwerk, M., Matsuoka, A., and Baumjohann, W.: Magnetometer in-flight offset accuracy for the BepiColombo spacecraft, *Ann. Geophys.*, 38, 823–832, 2020. <https://doi.org/10.5194/angeo-38-823-2020>
- Solomon, S. C., McNutt Jr., R. L., Gold, R. E., Domingue, D. L.: MESSENGER mission overview, *Space Sci. Rev.*, 131, 3–39, 2007. <https://doi.org/10.1007/s11214-007-9247-6>